

High resolution neutron radiography and microtomography with fast MCP-Timepix detector operating at >1Khz frame rates

Wednesday, 18 April 2012 11:30 (30 minutes)

The unique capability of MCP detectors to count neutrons with high detection efficiency (70% for cold and 50% for thermal), high spatial resolution (sub-15 μm) and no readout noise can be very attractive for some applications where relatively small area of the detector (currently $28 \times 28 \text{ mm}^2$) is acceptable. The recent development of fast parallel readout electronics for a 2×2 Timepix readout enabled high resolution imaging with event centroiding at \sim MHz counting rates. In addition to imaging resolution, the dynamical studies can be performed by time tagging of every neutron with accuracy of 1 μs (for 55 μm spatial resolution) and sub-ms for sub-15 μm spatial resolution. Another advantage of MCP detectors is their high dynamic range, allowing simultaneous event detection in case of very low neutron fluxes ($< 10 \text{ n/cm}^2/\text{s}$) and fluxes as high as $10^8 \text{ n/cm}^2/\text{s}$ with 1 KHz frames/s readout rates.

Our latest imaging experiments conducted at ICON and BOA beamlines will be presented, demonstrating not only the capabilities of the detection system, but also the excellent quality of those beamlines for high resolution neutron radiography and tomography. We will also describe the performance of the latest micropore neutron collimators used for scatter rejection in neutron radiography of samples with considerable neutron scattering, e.g. quantification of water content at a close distance to the detector required for high resolution imaging.

Primary author: Dr TREMSIN, Anton (University of California at Berkeley)

Co-authors: Dr KAESTNER, Anders (Paul Scherrer Institut); Dr LEHMANN, Eberhard (Paul Scherrer Institut); Dr MCPHATE, Jason (University of California at Berkeley); Dr VALERGA, John (University of California at Berkeley); Dr SIEGMUND, Oswald (University of California at Berkeley); Dr FELLER, W. Bruce (Nova Scientific, Inc.)

Presenter: Dr TREMSIN, Anton (University of California at Berkeley)

Session Classification: Water transport II