



Contribution ID: 8

Type: Poster

## Pulsed magnetic field measurement of the Nonlinear kicker for Korea-4GSR

The multipurpose synchrotron radiation accelerator (4GSR), supported by Pohang accelerator laboratory (PAL), aims to achieve an ultralow emittance of less than  $100 \text{ pm}\cdot\text{rad}$ , which requires a significantly reduced dynamic aperture size. The existing local bump injection scheme, which uses four kicker magnets to inject the beam, is unsuitable for 4GSR because it demands a large dynamic aperture. To solve this problem, PAL is researching the implementation of an injection method utilizing a single nonlinear kicker (NLK), with a prototype NLK made of G10 material. The magnetic field measurement system integrates the voltage across a single long coil to measure the magnetic field. It has an accuracy of less than  $10 \text{ }\mu\text{m}$  on each axis and can measure a space of  $470 \text{ mm} \times 50 \text{ mm} \times 750 \text{ mm}$ . The pulsed power supply is constructed with solid-state switches, capable of supplying a maximum pulse current of  $10 \text{ kA}$  for  $7 \text{ }\mu\text{s}$ .

**Primary authors:** Mr KIM, Beom Jun (Pohang Accelerator Laboratory); KIM, Dong Eon (Pohang Accelerator Laboratory (PAL)); JUNG, Young Gyu (Pohang Accelerator Laboratory (PAL)); LEE, Woulwoo (Pohang Accelerator Laboratory (PAL)); CHOI, Yoongeol (Pohang Accelerator Laboratory (PAL)); KIM, Min Jae (Pohang Accelerator Laboratory (PAL)); JUNG, Seong Hun (Pohang Accelerator Laboratory (PAL)); AN, Suk Ho (Pohang Accelerator Laboratory)

**Presenter:** Mr KIM, Beom Jun (Pohang Accelerator Laboratory)